



CTV PROTOCOL

Kush Desai
Chicago, IL

CT Venography of the abdomen/pelvis

- 125 mL contrast volume, injected at 3 mL/s, 50 mL saline chase
 - 130 second acquisition delay following injection
 - 5 mm slice thickness
 - Standard image reformats
-